

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/17/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:41
INTERNAL STANDARD	VERIFIED	02:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:43
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:44
BLANK TEST	0.00	02:44



505 Rep of Bartlett WESO #167 2/17/10