

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/25/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:28
INTERNAL STANDARD	VERIFIED	03:29
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:29
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:30
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:31
BLANK TEST	0.00	03:32



505 *Ralph Bartlett* 2/25/10 WESD #162