

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/19/2010

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:58
INTERNAL STANDARD	VERIFIED	14:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:58
BLANK TEST	0.00	14:59
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	14:59
BLANK TEST	0.00	15:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:00
BLANK TEST	0.00	15:01

J. M. Lopez
299
3/19/10

