

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 03/25/2010

Operator Name: RALPJ J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:53
INTERNAL STANDARD	VERIFIED	01:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:54
BLANK TEST	0.00	01:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:55
BLANK TEST	0.00	01:56



SGT Ralpj J Bartlett WCSO #162 03-25-10