

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/31/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:56
INTERNAL STANDARD	VERIFIED	01:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:56
BLANK TEST	0.00	01:57
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:57
BLANK TEST	0.00	01:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:58
BLANK TEST	0.00	01:59



505 *Ralph Bartlett* WESO #162

03/31/10