

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206




Date: 04/23/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:29
INTERNAL STANDARD	VERIFIED	00:29
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:29
BLANK TEST	0.00	00:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:30
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:32

⊙Ç3W▼≈



505 Ralph Bartlett      WSP #162      4/23/10