

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/02/2010

Operator Name: CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:14
INTERNAL STANDARD	VERIFIED	08:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	08:14
BLANK TEST	0.00	08:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:15
BLANK TEST	0.00	08:16
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	08:16
BLANK TEST	0.00	08:17

*Charles B. Geisler*

