

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/19/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:50
INTERNAL STANDARD	VERIFIED	15:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:50
BLANK TEST	0.00	15:51
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:51
BLANK TEST	0.00	15:52
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:52
BLANK TEST	0.00	15:53



*SOB Rusty Bartlett 5/19/10 wes #162*