

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/27/2010

Operator Name: CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:41
INTERNAL STANDARD	VERIFIED	09:41
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	09:41
BLANK TEST	0.00	09:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:42
BLANK TEST	0.00	09:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:43
BLANK TEST	0.00	09:44

*So. Ch. B. Geisler #248*

