

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/01/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:28
INTERNAL STANDARD	VERIFIED	10:28
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	10:29
BLANK TEST	0.00	10:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:30
BLANK TEST	0.00	10:30
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	10:31
BLANK TEST	0.00	10:31



Charles B. Geisler #248 6/1/10 1032