

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/18/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:53
INTERNAL STANDARD	VERIFIED	03:53
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:54
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	03:54
BLANK TEST	0.00	03:55
REFERENCE STANDARD TEMP	34.05c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:56



565 Ralph Bartlett

06/18/10