

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 06/27/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:23
INTERNAL STANDARD	VERIFIED	01:23
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:26



SCS Ralph Bartlett

06/27/10