

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 07/01/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:10
INTERNAL STANDARD	VERIFIED	04:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:10
BLANK TEST	0.00	04:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:12
BLANK TEST	0.00	04:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:13
BLANK TEST	0.00	04:13



SGT Rusty Burt

07/01/10  
07/01/10