

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 07/07/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:50
INTERNAL STANDARD	VERIFIED	01:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:53



565 Rusty Jones

7/7/10