

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 07/15/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:03
INTERNAL STANDARD	VERIFIED	19:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:03
BLANK TEST	0.00	19:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:04
BLANK TEST	0.00	19:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:05
BLANK TEST	0.00	19:06



SGT Ralph J Bartlett

WCSO #102

07-15-10