

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 07/22/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:08
INTERNAL STANDARD	VERIFIED	00:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:09
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:10
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12



SG5 Ralph J Bartlett 07-22-10 WCRS #162