

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 08/09/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:34
INTERNAL STANDARD	VERIFIED	03:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:36
BLANK TEST	0.00	03:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:37
BLANK TEST	0.00	03:37



87 *Ralph J Bartlett* WSO #162 8/9/10