

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 08/11/2010

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:31
INTERNAL STANDARD	VERIFIED	16:31
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	16:31
BLANK TEST	0.00	16:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:32
BLANK TEST	0.00	16:33
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	16:33
BLANK TEST	0.00	16:34

*J.M. Lopez #3799*

