

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 08/23/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:03
INTERNAL STANDARD	VERIFIED	04:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:04
BLANK TEST	0.00	04:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:05
BLANK TEST	0.00	04:06



SOT Ralph J Bartlett 08-23-10

WCSO #102