

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 08/31/2010

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:17
INTERNAL STANDARD	VERIFIED	01:17
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:20



565 *Ralph Bartlett* 08-31-10 WCO Sheriff #162