

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 09/06/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:08
INTERNAL STANDARD	VERIFIED	05:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	05:09
BLANK TEST	0.00	05:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:10
BLANK TEST	0.00	05:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:11
BLANK TEST	0.00	05:12



505 *Ralph Bartlett*

WESO #162

09-06-10