

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206




Date: 09/14/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10110

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 04:19 |
| INTERNAL STANDARD | VERIFIED | 04:19 |
| REFERENCE STANDARD TEMP | 33.97c | |
| REFERENCE STANDARD | 0.09 | 04:20 |
| BLANK TEST | 0.00 | 04:20 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.09 | 04:21 |
| BLANK TEST | 0.00 | 04:21 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.09 | 04:22 |
| BLANK TEST | 0.00 | 04:23 |

||-b8W▼ëb■°



505 *Ralph Bartlett*

WCSO #162

09-14-10