

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 09/20/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:51
INTERNAL STANDARD	VERIFIED	02:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:52
BLANK TEST	0.00	02:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55



565 *Ralph Bartlett* 09/20/10 meso 51410 #162