

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/07/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:43
INTERNAL STANDARD	VERIFIED	03:44
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:44
BLANK TEST	0.00	03:45
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:45
BLANK TEST	0.00	03:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:46
BLANK TEST	0.00	03:47



see *Ralph Bartlett* wsc Sheriff #162 10/7/10