

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 10/13/2010

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:46
INTERNAL STANDARD	VERIFIED	01:46
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:46
BLANK TEST	0.00	01:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:48
BLANK TEST	0.00	01:49



SGS Rugby Bureau

10/12/10

WCSO #162