

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/21/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:54
INTERNAL STANDARD	VERIFIED	04:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:55
BLANK TEST	0.00	04:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:56
BLANK TEST	0.00	04:56
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	04:57
BLANK TEST	0.00	04:58



See Ralph Bartlett 10/21/10 WOSO #162