

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/27/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:15
INTERNAL STANDARD	VERIFIED	03:15
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:16
BLANK TEST	0.00	03:16
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:17
BLANK TEST	0.00	03:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:18
BLANK TEST	0.00	03:19



SUT Ralph Bartlett

WSP #162

10/27/10