

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 11/01/2010

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:08
INTERNAL STANDARD	VERIFIED	04:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:08
BLANK TEST	0.00	04:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:09
BLANK TEST	0.00	04:10
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	04:10
BLANK TEST	0.00	04:11



SGT Ralph J Bartlett WOSO SHERIFF #162 11-01-10