

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 11/09/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:05
INTERNAL STANDARD	VERIFIED	03:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:06
BLANK TEST	0.00	03:07
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:08
BLANK TEST	0.00	03:09



SGT *Ralph Bartlett*

WCSO #162

11-09-10