

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 11/24/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:32
INTERNAL STANDARD	VERIFIED	09:32
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	09:33
BLANK TEST	0.00	09:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:34
BLANK TEST	0.00	09:35
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	09:35
BLANK TEST	0.00	09:36

Charles B. Geisler #248

