SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/03/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:02
INTERNAL STANDARD	VERIFIED	12:02
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	12:03
BLANK TEST	0.00	12:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:04
BLANK TEST	0.00	12:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:05
BLANK TEST	0.00	12:06







#248