

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 12/26/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:27
INTERNAL STANDARD	VERIFIED	01:27
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:29
BLANK TEST	0.00	01:30

