

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/04/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:29
INTERNAL STANDARD	VERIFIED	03:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:30
BLANK TEST	0.00	03:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:31
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33



SOS Ralph Bartlett

West SHERID #162

01/04/11