

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/10/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:38
INTERNAL STANDARD	VERIFIED	03:38
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:40
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:41
BLANK TEST	0.00	03:42



SGT Ralph Bartlett used #12 1/10/11