

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/18/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:32
INTERNAL STANDARD	VERIFIED	03:32
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:33
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:35
BLANK TEST	0.00	03:36



SGT *Ralph Bartlett* WESO #162 1/18/11