

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/01/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:23
INTERNAL STANDARD	VERIFIED	02:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:25
BLANK TEST	0.00	02:26



505 High Purity NCSO #162

2/1/11