

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 02/15/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:47
INTERNAL STANDARD	VERIFIED	03:47
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:48
BLANK TEST	0.00	03:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:49
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	03:50
BLANK TEST	0.00	03:50



See Ralph Bartlett were #162 02-15-11