

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/21/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:35



SGT Ralph Bartlett

WCSO #162

02-21-11