

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 03/21/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:04
INTERNAL STANDARD	VERIFIED	04:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:04
BLANK TEST	0.00	04:05
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:05
BLANK TEST	0.00	04:06
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:06
BLANK TEST	0.00	04:07



SOS Ralph J Bartlett WOSO #162 3/21/11