

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 03/30/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:46
INTERNAL STANDARD	VERIFIED	02:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:47
BLANK TEST	0.00	02:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:49



305 High Street WOSO Sheriff #162 3/30/11