

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 04/07/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:10
INTERNAL STANDARD	VERIFIED	04:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:11
BLANK TEST	0.00	04:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:12
BLANK TEST	0.00	04:13
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:13
BLANK TEST	0.00	04:14



565 Ralph J Bartlett

WESO #162

04/07/11