

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 04/21/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:32
INTERNAL STANDARD	VERIFIED	02:32
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:35
BLANK TEST	0.00	02:36



*SGT Ralph Bartlett*

04-21-11

WASO #162