

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/02/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:57
INTERNAL STANDARD	VERIFIED	04:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:58
BLANK TEST	0.00	04:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:59
BLANK TEST	0.00	05:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	05:00
BLANK TEST	0.00	05:01



SGT *Ralph Bartlett* WCSO #162 5/2/11