

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/11/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:46
INTERNAL STANDARD	VERIFIED	03:46
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	03:46
BLANK TEST	0.00	03:47
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:47
BLANK TEST	0.00	03:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:48
BLANK TEST	0.00	03:49



505 *Ralph J Bartlett*

5/11/11

WCSO #162