

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/16/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:45
INTERNAL STANDARD	VERIFIED	02:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:46
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:47
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:48



SOE Ralph J Bartlett WESO #162 05/16/11