

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/08/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:34
INTERNAL STANDARD	VERIFIED	03:34
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:35
BLANK TEST	0.00	03:36
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:37
BLANK TEST	0.00	03:37



SGT Ralph J Bartlett WCSO #162 06-08-11