

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/22/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:14
INTERNAL STANDARD	VERIFIED	03:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:14
BLANK TEST	0.00	03:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:15
BLANK TEST	0.00	03:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:16
BLANK TEST	0.00	03:17



Sgt. Ralph Bartlett

WARREN CO. SO

06-22-11