

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/30/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:38
INTERNAL STANDARD	VERIFIED	03:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:40
BLANK TEST	0.00	03:41



565 Ralph Bartlett WESO #162 06-30-11