

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 07/14/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:54
INTERNAL STANDARD	VERIFIED	03:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:55
BLANK TEST	0.00	03:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:57
BLANK TEST	0.00	03:58



SGT *Ralph Bartlett* WESO #112 07-14-11