

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/02/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:43
INTERNAL STANDARD	VERIFIED	03:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:44
BLANK TEST	0.00	03:44
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:45
BLANK TEST	0.00	03:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:46
BLANK TEST	0.00	03:47



565 *Ralph Bartlett*

WCSO #162

08-02-11